

Certificate of Analysis: S-316 Oil Extraction Solvent

Lot Number: 8C20146

Manufacturer: HORIBA

CAS Number

9002-83-9 100%

Analysis	Specification	Result
Appearance	Clear and Colorless	Pass
FTIR Scan Wavenumber 4000-2800 cm ⁻¹	< 0.05 Absorbance	Pass
Karl Fisher Water	< 55 ppm	Pass
Acid Number	< 0002 mg KOH/g	Pass
Separation Time	< 15 seconds	Pass
Viscosity @ 100°F	0.72 – 1.1 cs	Pass

Analysis Date: 08/12/2020

This is to certify that HORIBA S-316 extraction solvent shipped with this lot number conforms to the material specifications established by HORIBA Instruments, Incorporated. This certificate is valid only if the extraction solvent is not mishandled, contaminated or modified. The S-316 solvent should be opened for the minimum time required to dispense the solvent.

Phone: 800-446-7422 or 949-250-4811

FAX: 949-250-0924

www.horiba-water.com